



Our STN: BL 125664/2

**SUPPLEMENT APPROVAL**

November 7, 2019

Siwa Biotech Corp  
Attention: William Canfield, MD, PhD  
755 Research Parkway, Suite 125  
Oklahoma City, OK 73104-3631

Dear Dr. Canfield:

We have approved your request submitted August 12, 2019, received August 15, 2019, to supplement your Biologics License Application (BLA) under section 351(a) of the Public Health Service Act for Blood Grouping Reagent, Anti-Fy<sup>a</sup> (Murine Monoclonal) (Recombinant), product code QHR. You requested revisions to the Instructions for Use to include a 2-minute post-centrifugation read time and a limitations statement regarding weakened or false negative reactions if prompt centrifugation is not performed after the addition of red cell suspension.

We hereby approve the draft package insert labeling submitted on August 12, 2019. This is a reminder that as of September 24, 2014, medical devices that are licensed under the PHS Act are subject to certain provisions of the final Unique Device Identifier (UDI) rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR 830 Subpart E). Additionally, please identify each device identifier implemented for the subject device, and the device identifiers that have been discontinued for the subject device as a labeling change in an annual report consistent with 21 CFR 601.12(f)(3). For more information on these requirements, please see the UDI website at <http://www.fda.gov/udi>.

Please submit all final printed labeling as PDF electronic copy (eCopy) at the time of use and include implementation information on Form FDA 356h as appropriate.

Two draft copies of the proposed introductory advertising or promotional labeling may be voluntarily submitted for advisory comment with a completed Form FDA 2253 to the Advertising and Promotional Labeling Branch at the following address:

Food and Drug Administration  
Center for Biologics Evaluation and Research  
Document Control Center  
10903 New Hampshire Ave.  
WO71-G112  
Silver Spring, MD 20993-0002

Please submit an amendment to all pending supplemental applications for this BLA that include revised labeling incorporating a revised content of labeling that includes these changes.

We will include information contained in the above-referenced supplement in your BLA file.

Sincerely,

Orieji Illoh, MD  
Director  
Division of Blood Components and Devices  
Office of Blood Research and Review  
Center for Biologics Evaluation and Research