

# BLOOD GROUPING REAGENT

**Anti-Jk<sup>b</sup>**  
**ALBAclone<sup>®</sup>**  
**(Monoclonal)**

**For Tube Techniques**

**REF Z166U**

- **FOR *IN VITRO* DIAGNOSTIC USE**
- **Meets FDA potency requirements**
- **Discard if turbid**
- **Preservative: 0.09% (w/v) sodium azide and 0.02% (w/v) sodium arsenite**

CAUTIONS: THE ABSENCE OF ALL VIRUSES HAS NOT BEEN DETERMINED. THIS PRODUCT HAS COMPONENTS (DROPPER BULBS) CONTAINING DRY NATURAL RUBBER.

## INTERPRETATION OF LABELING SYMBOLS

LOT

Batch code



Use by (YYYY-MM-DD)

REF

Product code



Storage temperature limitation (2-8 °C)

IVD

*In vitro* diagnostic medical device



Consult instructions for use

[www.quotientbd.com](http://www.quotientbd.com)



Manufacturer

## INTENDED USE

This Anti-Jk<sup>b</sup> reagent is for the *in vitro* detection and identification of the human Jk<sup>b</sup> blood group antigen by direct agglutination.

## SUMMARY AND EXPLANATION

Anti-Jk<sup>a</sup> and Anti-Jk<sup>b</sup> (Anti-JK1 and Anti-JK2) were described in 1951 and 1953 respectively and define a pair of alleles on the long arm of chromosome 18.

The Kidd system is particularly important in clinical practice. Anti-Jk<sup>a</sup> and anti-Jk<sup>b</sup> in patient samples often demonstrate an inherent lack of stability and an inability to agglutinate cells which express a single dose of antigen. Kidd antibodies have been implicated in cases of delayed hemolytic transfusion reactions.

## PRINCIPLE OF THE TEST

When used by the recommended technique, this reagent will cause the agglutination (clumping) of red blood cells carrying the Jk<sup>b</sup> antigen. Lack of agglutination demonstrates the absence of the Jk<sup>b</sup> antigen.

## REAGENT DESCRIPTION

The main component of this reagent is derived from the *in vitro* culture of the IgM secreting human/mouse heterohybridoma:

Product Name	Product Code	Cell Line
Anti-Jk <sup>b</sup>	Z166U	P3.143

The formulation also contains 0.09% (w/v) sodium azide and 0.02% (w/v) sodium arsenite.

NOTE: The volume delivered by the reagent bottle dropper is approximately 40 µL. Care should be taken to ensure that appropriate serum to cell ratios are maintained in all test systems.

## WARNINGS AND PRECAUTIONS

For *in vitro* diagnostic use only  
Products should be used by qualified personnel  
Do not use beyond the expiration date  
Do not use if turbid  
Do not dilute

The format of the expiration date is expressed as YYYY-MM-DD (Year-Month-Day)

This reagent contains 0.09% (w/v) sodium azide. Sodium azide may be toxic if ingested and may react with lead and copper plumbing to form explosive compounds. If discarded into a sink, flush with a large volume of water to prevent azide build-up.

This reagent contains material of animal origin (murine and bovine), therefore care must be taken during use and disposal as there is a potential infection risk.

CAUTION: ALL BLOOD PRODUCTS SHOULD BE TREATED AS POTENTIALLY INFECTIOUS. SOURCE MATERIAL FROM WHICH THIS PRODUCT WAS DERIVED WAS FOUND NEGATIVE WHEN TESTED IN ACCORDANCE WITH CURRENT FDA REQUIRED TESTS. NO KNOWN TEST METHODS CAN OFFER ASSURANCE THAT PRODUCTS DERIVED FROM HUMAN BLOOD WILL NOT TRANSMIT INFECTIOUS AGENTS.

The bovine material used in the manufacture of this reagent was collected in a USDA approved facility or obtained from a geographical region classified as negligible risk for BSE.

Monoclonal antibodies exhibit a high degree of potency, avidity and specificity. When using such antibodies, great care should be taken to avoid cross contamination. This product has components (dropper bulbs) containing dry natural rubber.

## STORAGE

The reagent should be stored at 2-8 °C.

## SPECIMEN COLLECTION AND PREPARATION

Specimens should be collected by a standard collection technique. The specimen should be tested as soon as possible after collection. If testing is delayed, the specimen should be stored at refrigerated temperatures.

Clotted samples, or those collected in EDTA, should be tested within fourteen days from collection. Donor blood collected in ACD, CPD, CPDA -1, CP2D, CP2D with AS-3, CPD with AS-1, and CPD with AS-5 may be tested until the expiration date of the donation.

Special care should be taken if hemolyzed samples must be tested. Grossly icteric or contaminated blood specimens should not be used.

## MATERIALS

### Material provided

- ALBAclone<sup>®</sup> Anti-Jk<sup>b</sup>

### Materials required but not provided

- Isotonic saline
- Reagent red blood cells suitable for the control of Anti-Jk<sup>b</sup>
- 10 x 75 mm or 12 x 75 mm glass test tubes
- Pipettes
- Optical aid (optional)
- Centrifuge

## PROCEDURE

NOTE: This reagent has been standardized for use by the technique described below and therefore its suitability for use by other techniques cannot be guaranteed. When a test is required to be incubated for a specific time period, a timer should be used.

When using supplemental testing equipment (i.e. centrifuge), follow the procedures that are contained in the operator's manual provided by the device manufacturer.

### Tube Technique - Immediate Spin

1. Prepare a 2-4% suspension of red blood cells in isotonic saline solution (Reagent Red Blood Cells may be used directly from the vial or according to the manufacturer's instructions.)
2. Add 1 drop of blood grouping reagent to a glass test tube.
3. Add 1 drop of red blood cell suspension. Steps 2 and 3 may be performed in either order.
4. Mix the contents of the test tube and centrifuge.

NOTE: Suggested centrifugation: 900-1000 g (approx. 3400 rpm) for 10 seconds or a time and speed appropriate for the centrifuge used that produces the strongest reaction of

antibody with antigen-positive red blood cells, yet allows easy re-suspension of antigen-negative red blood cells.

- After centrifugation, gently shake the tube to dislodge the cell button from the bottom and immediately observe macroscopically for agglutination. Negative reactions may be examined with an optical aid.
- Record results.

**Refer to Performance Limitations section for additional guidance on the use of this product**

**STABILITY OF REACTION**

Test results should be read, interpreted and recorded immediately after centrifugation. Delays may cause dissociation of antigen-antibody complexes resulting in weak positive or false negative reactions.

**INTERPRETATION OF RESULTS**

Agglutination = positive test result  
 No agglutination = negative test result

**QUALITY CONTROL**

Quality control of reagents is essential and should be performed on each day of use and in accordance with local, state and federal regulations.

JK(a+b+) red blood cells should be used as a positive control  
 JK(a+b-) red blood cells should be used as a negative control

Other red blood cell types may be suitable but should be selected with care.

False positive test results are rarely seen with low-protein reagents. False positive agglutination may be due to a positive direct antiglobulin test (DAT), cold agglutinins, or abnormal serum proteins. If false positive results are suspected, or local regulations require, and a control test for spontaneous agglutination is desired, ALBAcheck® - BGS Monoclonal Control (Z271U) or 6-10% albumin in saline may be substituted for the blood grouping reagent in the testing procedure. A negative result would serve as an appropriate control. If the monoclonal control test gives a positive reaction, a valid interpretation of the results obtained in red blood cell testing cannot be made without further investigation.

**PERFORMANCE LIMITATIONS**

The expression of certain red blood cell antigens may diminish in strength during storage, particularly in EDTA and clotted samples. Better results will be obtained with fresh samples.

Gently re-suspend tube tests before reading. Excessive agitation may disrupt weak agglutination and produce false negative results.

Excessive centrifugation can lead to difficulty in re-suspending the cell button, while inadequate centrifugation may result in agglutinates that are easily dispersed.

False positive or false negative results can occur due to contamination of test materials, improper reaction temperature,

improper storage of materials, omission of test reagents and certain disease states.

Suppressed or weak expression of blood group antigens may give rise to false negative reactions.

Care should be taken when testing red blood cells that have been treated with proteolytic enzymes, as these may produce false positive or false negative results.

**SPECIFIC PERFORMANCE CHARACTERISTICS**

Prior to release, each lot of ALBAclone® Anti-Jk<sup>b</sup> is tested using FDA recommended methods against a panel of antigen-positive and antigen-negative red blood cells to ensure suitable reactivity.

In performance evaluation studies (data on file at Alba Bioscience Limited), blood group samples were tested with ALBAclone® Anti-Jk<sup>b</sup> as follows:

Reagent	No. Samples Tested	Concordance*
ALBAclone® Anti-Jk <sup>b</sup>	1108	100%

\* Concordance indicated agreement between the ALBAclone® Anti-Jk<sup>b</sup> and comparator reagents only and does not indicate which reagent gave the correct results.

Repeatability and reproducibility of the trial reagent was confirmed by means of Lot to Lot and Precision studies.

**Comparator Study Results**

During comparator studies (data on file at Alba Bioscience Limited), blood samples were tested with ALBAclone® Anti-Jk<sup>b</sup> (Monoclonal) (IgM) as follows:

Anti-Jk <sup>b</sup>		Comparator Reagent			One-sided 95% Exact lower confidence limit
		Positive	Negative	Total	
Trial Reagent	Positive	757	0	757	
	Negative	0	351	351	
	Total	757	351	1108	
<b>Positive Percent Agreement*</b>				100	99.61%
<b>Negative Percent Agreement*</b>				100	99.15%

\* Indicates agreement between the ALBAclone® Anti-Jk<sup>b</sup> and comparator reagents only and does not indicate which reagent gave the correct result(s).

In performance evaluation studies, 1108 samples were tested with ALBAclone® Anti-Jk<sup>b</sup> (Monoclonal) (IgM). The positive percent agreement at the one-sided 95% exact lower confidence limit was 99.61% for agglutination tests based on a comparison of interpreted results. The negative agreement at the one-sided 95% exact lower confidence limit was 99.15% for agglutination tests based on a comparison of interpreted results.

Results were evaluated against comparable FDA approved products using the appropriate methods for the comparators.

**Precision Study Results**

As part of the performance evaluation, precision and lot to lot studies were performed using multiple operators, days and runs to confirm repeatability and reproducibility of test results in the same run, day and with the same operator and between runs, days and operators. The study took account of variables such as days of the week, times of day and supplementary reagents used in the testing.

All antigen positive test outcomes generated unequivocal positive reactions and antigen negative test outcomes generated unequivocal negative reactions.

**BIBLIOGRAPHY**

- Roback JD, Grossman BJ, Harris T, *et al*: AABB Technical Manual, 18<sup>th</sup> ed. AABB, 2014
- Daniels G: Human Blood Groups, 3<sup>rd</sup> ed. Blackwell Science, 2012
- AABB Standards Program Committee: Standards for Blood Banks and Transfusion Services, 29<sup>th</sup> ed. AABB, 2014
- Reid ME, Lomas-Francis C, Olsson ML: The Blood Group Antigen FactsBook, ed 3. Academic Press, 2012
- Klein HG, Anstee DJ: Mollison's Blood Transfusion in Clinical Medicine, 11<sup>th</sup> ed. Blackwell Publishing, 2005
- Daniels G, Bromilow I: Essential Guide to Blood Groups, 2<sup>nd</sup> ed. Blackwell Publishing Ltd, 2012

**DATE OF ISSUE**

YYYY-MM-DD

US Distributor

Quotient  
 301 South State Street  
 S-204  
 Newtown  
 PA 18940  
 USA

Customer Service Tel: 1-888-284-1901  
 Product Technical Support Tel: 1-888-228-1990  
 Customer Service Fax: 1-888-694-5208  
 E-Mail: customer.serviceUS@quotientbd.com  
 Web: www.quotientbd.com/us



Alba Bioscience Limited  
 James Hamilton Way  
 Penicuik  
 UK  
 EH26 0BF  
 US License 1807