



September 3, 2021

Immucor, Inc.
Attention: Mr. Howard Yorek
3130 Gateway Drive
Norcross, GA 30071

Re:

STN	Trade/Device Name
BK210600	NEO Iris
BK210605	Galileo NEO

Regulation Number: 21 CFR 864.9175
Regulation Name: Automated blood grouping and antibody test system
Regulatory Class: Class II
Product Code: KSZ
Dated: June 9, 2021
Received: June 14, 2021

Dear Mr. Yorek:

We have reviewed your Section 510(k) premarket notification of intent to market the devices referenced above and have determined the devices are substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your products as devices, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your devices are classified (see above) into either class II (Special Controls) or class III (PMA), they may be subject to additional controls. Existing major regulations affecting your devices can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In

addition, FDA may publish further announcements concerning your devices in the **Federal Register**.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your devices comply with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Orieji Illoh, MD
Director
Division of Blood Components and Devices
Office of Blood Research and Review
Center for Biologics Evaluation and Research

Enclosure
Indications for Use

Indications for Use

510(k) Number: BK210600 and BK210605

Device Name: NEO Iris (BK210600)
Galileo NEO (BK210605)

Indications for Use:

The NEO Iris is a microprocessor-controlled instrument to fully automate immunohematology in vitro diagnostic testing of human blood. The NEO Iris automates test processing, result interpretation and data management functions. The NEO Iris is designed to automate standard immunohematology assays using a microplate-based platform. Assays include ABO grouping and Rh (D) typing, detection/identification of IgG red blood cell antibodies, compatibility testing, red blood cell phenotyping and antigen screening.

The NEO Iris is for in vitro diagnostic use.

The Galileo NEO is a microprocessor-controlled instrument to fully automate immunohematology in vitro diagnostic testing of human blood. The Galileo NEO automates test processing, result interpretation and data management functions. The Galileo NEO is designed to automate standard immunohematology assays using a microplate-based platform. Assays include ABO grouping and Rh (D) typing, detection/identification of IgG red blood cell antibodies, compatibility testing, red blood cell phenotyping and antigen screening.

The Galileo NEO is for in vitro diagnostic use.

Prescription Use AND/OR Over-The-Counter Use
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CBER, Office of Blood Research and Review

Division Sign-Off
Office of Blood Research and Review