



June 16, 2023

Jo-Ann F. Gonzales, RAC  
Director, IVD Regulatory Consulting  
Precision for Medicine  
Representing:  
PHASE Scientific International, Ltd.  
10527 Garden Grove Blvd  
Garden Grove, CA, 92843

Re: EUA220152/S008  
Trade/Device Name: INDICAID COVID-19 Rapid Antigen At-Home Test  
Dated: April 27, 2023  
Received: April 27, 2023

Dear Jo-Ann Gonzales:

This is to notify you that your request to update the INDICAID COVID-19 Rapid Antigen At-Home Test to extend the shelf-life expiration date to 15 months when stored at 2 – 30°C based on the results of your extended stability studies, is granted. Upon review, we concur that the data and information submitted in EUA220152/S008 supports the requested updates for use with the INDICAID COVID-19 Rapid Antigen At-Home Test. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the INDICAID COVID-19 Rapid Antigen At-Home Test reissued on March 16, 2022.

Sincerely yours,

---

Uwe Scherf, M.Sc., Ph.D.  
Director, Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health