

Contains Nonbinding Recommendations

Draft – Not for Implementation

Recommendations for the Use of Clinical Data in Premarket Notification [510(k)] Submissions

Draft Guidance for Industry and Food and Drug Administration Staff

DRAFT GUIDANCE

This draft guidance document is being distributed for comment purposes only.

Document issued on September 7, 2023.

You should submit comments and suggestions regarding this draft document within 90 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD 20852-1740. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions about this document regarding CDRH-regulated devices, contact the Office of Product Evaluation and Quality, Office of Regulatory Programs, Division of Regulatory Programs 1 (Submission Support) at 301-796-5640. For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach, and Development (OCOD) at 1-800-835-4709 or 240-402-8010, or by email at ocod@fda.hhs.gov.



U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Center for Biologics Evaluation and Research

Contains Nonbinding Recommendations

Draft – Not for Implementation

Preface

Additional Copies

CDRH

Additional copies are available from the Internet. You may also send an email request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please include the document number GUI00020016 and complete title of the guidance in the request.

CBER

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-8010, by email, ocod@fda.hhs.gov, or from the Internet at <https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

Table of Contents

I. Introduction..... 1

II. Background..... 2

III. Scope..... 5

IV. Appropriate Use of Clinical Data in 510(k) Decision-Making..... 6

V. Scenarios When Clinical Data May be Necessary to Determine Substantial Equivalence 6

 A. Scenario #1 – Differences in the indications for use..... 7

 B. Scenario #2 – Differences in the technological characteristics..... 8

 C. Scenario #3 – SE cannot be determined by non-clinical testing..... 10

 D. Scenario #4 – Newly identified or increased risk for the predicate device..... 11

Recommendations for the Use of Clinical Data in Premarket Notification [510(k)] Submissions

Draft Guidance for Industry and Food and Drug Administration Staff

This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.

I. Introduction

As part of FDA’s [Medical Device Safety Action Plan: Protecting Patients, Promoting Public Health](#) (herein referred to as the “Safety Action Plan”),¹ FDA committed to strengthen and modernize the premarket notification [510(k)] Program. FDA is issuing this guidance to provide our current thinking on the use of clinical data in 510(k) submissions to enhance the predictability, consistency, and transparency of the 510(k) Program. The intent of this guidance is to clarify and provide additional context for situations when clinical data may be necessary to demonstrate substantial equivalence (SE), as initially described in “[The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \[510\(k\)\]](#)” guidance (herein referred to as the “510(k) Program Guidance”).²

In that guidance, FDA described the most common scenarios for when clinical data may be necessary in a 510(k) submission. The scenarios are further described in this guidance, and FDA has described another scenario. In addition, FDA is providing additional examples to clarify these concepts, illustrating when clinical data may or may not be needed. Providing clarity and predictability about when clinical data may be necessary to include in a 510(k) submission to demonstrate SE will aid in protecting and promoting public health.

¹ Available at <https://www.fda.gov/about-fda/cdrh-reports/medical-device-safety-action-plan-protecting-patients-promoting-public-health>.

² Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

33 In general, FDA’s guidance documents do not establish legally enforceable responsibilities.
34 Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only
35 as recommendations, unless specific regulatory or statutory requirements are cited. The use of
36 the word *should* in Agency guidances means that something is suggested or recommended, but
37 not required.
38

39 **II. Background**

40 In April 2018, CDRH issued the [Safety Action Plan](#) to communicate CDRH’s vision for
41 modernizing measures to improve the safety of medical devices while continuing to create more
42 efficient pathways to bring critical devices to patients. The [Safety Action Plan](#) describes efforts
43 underway to enhance our programs to help improve device safety.
44

45 In November 2018, FDA announced transformative additional steps to modernize FDA’s 510(k)
46 Program to advance the review of the safety and effectiveness of medical devices.³ In connection
47 with this announcement, FDA also requested public feedback on these steps to continue to
48 modernize the framework for 510(k) review while promoting patient safety and posed other
49 questions that could inform regulatory policy development.⁴ One area identified by the public
50 comments where additional clarity and transparency would be helpful was the use of clinical data
51 in 510(k) submissions.
52

53 Under section 510(k) of the Federal Food, Drug, and Cosmetic (FD&C) Act, a premarket
54 notification submission (often referred to as a 510(k)) must be submitted to FDA at least 90 days
55 before introducing, or delivering for introduction, a device into interstate commerce for
56 commercial distribution.⁵ A 510(k) is required for devices intended for human use, for which a
57 premarket approval application (PMA) is not required, unless the device is exempt from the
58 510(k) requirements of the FD&C Act and does not exceed the relevant limitations of
59 exemptions in the device classification regulations. Through review of the 510(k), FDA
60 determines whether the “new device”⁶ is substantially equivalent⁷ (SE) to a predicate device.⁸

³ Please see the Statement from then FDA Commissioner Scott Gottlieb, M.D., and Jeff Shuren, M.D., Director of the Center for Devices and Radiological Health, on November 26, 2018, available at <https://www.fda.gov/news-events/press-announcements/statement-fda-commissioner-scott-gottlieb-md-and-jeff-shuren-md-director-center-devices-and>.

⁴ Please see “Modernizing FDA’s 510(k) Program; Establishment of a Public Docket; Request for Comments,” Docket Number FDA-2018-N-4751, available at <https://www.regulations.gov/docket/FDA-2018-N-4751>.

⁵ Under section 510(k) of the FD&C Act, a 510(k) is required for devices that are not subject to a premarket approval application, unless the device is exempt from the 510(k) requirements of the FD&C Act and does not exceed the limitations of exemptions for each of the device classification regulations (see 21 CFR Parts 862-892). See sections 510(k) and (n) of the FD&C Act (21 U.S.C. §§ 360(k) & (n)).

⁶ For purposes of this guidance, a “new device” means a device within the meaning of section 201(h) of the FD&C Act that is not legally marketed. It can be either a completely new device or a modification of a legally marketed device that would require a new 510(k).

⁷ The standard for a substantial equivalence determination for a 510(k) submission is set out in section 513(i) of the FD&C Act.

⁸ For purposes of an SE determination, a predicate device is (1) a device that was legally marketed prior to May 28, 1976 (preamendments device) and for which a PMA is not required, or (2) a device that has been classified or

Contains Nonbinding Recommendations

Draft – Not for Implementation

61 For additional information on how FDA evaluates SE in the 510(k) review process, please see
62 the [510\(k\) Program Guidance](#).

63
64 For FDA to find a new device SE to a predicate device, FDA must first find that the new device
65 and predicate device have the same intended use. FDA must then find that the new device and
66 predicate device have the same technological characteristics, or if they do not, that the different
67 technological characteristics⁹ of the new device do not raise different questions of safety and
68 effectiveness and that the new device is as safe and effective as a predicate device.

69
70 To determine the safety and effectiveness of a device, FDA also weighs if there is “any probable
71 benefit to health from the use of the device against any probable risk of injury or illness from
72 such use,”¹⁰ among other relevant factors. Under the 510(k) paradigm, the benefit-risk profile of
73 the new device is determined in the context of a comparison to the benefit-risk profile of a
74 predicate device; the benefit-risk profile of a new device with different technological
75 characteristics does not need to be identical to that of its predicate device in order to determine if
76 the new device is as safe and effective as a predicate device. The FDA guidance “[Benefit-Risk
77 Factors to Consider When Determining Substantial Equivalence in Premarket Notifications
78 \(510\(k\)\) with Different Technological Characteristics](#)”¹¹ describes considerations for evaluating
79 benefit-risk profile of a device in comparison to a predicate device for purposes of SE
80 determinations.

81
82 In many cases, a new device that is subject to 510(k) requirements can demonstrate SE to a
83 predicate device through robust non-clinical safety and performance data, without the need for
84 clinical data, for example, because the intended use and technological characteristics of the new
85 device is the same as, or sufficiently similar to, that of the predicate device. In such
86 circumstance, clinical data would not be necessary to demonstrate SE to a predicate device, and
87 requiring clinical data would be inconsistent with the least burdensome provisions of the FD&C
88 Act.¹² However, for certain devices subject to 510(k) requirements, obtaining clinical data may
89 be necessary to demonstrate that a new device is SE to a predicate device.

90
91 As described in the [510\(k\) Program Guidance](#), when analytical or non-clinical bench
92 performance testing data or non-clinical animal¹³ and/or biocompatibility studies are insufficient,

reclassified into Class II or I, or (3) a device that has been found to be SE through the 510(k) process. See 21 CFR 807.92(a)(3) and section 513(f)(2) of the FD&C Act.

⁹ A predicate device is a legally marketed device. For purposes of an SE determination, “different technological characteristics” means, with respect to a device being compared to a predicate device, that there is a significant change in the materials, design, energy source, or other features of the device from those of the predicate device.” See section 513(i)(1)(B) of the FD&C Act.

¹⁰ See 21 CFR 860.7(b).

¹¹ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/benefit-risk-factors-consider-when-determining-substantial-equivalence-premarket-notifications-510k>.

¹² See section 513(i)(1)(D)(ii) of the FD&C Act, and “[The Least Burdensome Provisions: Concept and Principles](#)” guidance, available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/least-burdensome-provisions-concept-and-principles>.

¹³ FDA supports the principles of the “3Rs” to replace, reduce, and/or refine animal use in testing, when feasible. We encourage manufacturers to consult with FDA if they wish to use a non-animal testing method that they believe

Contains Nonbinding Recommendations

Draft – Not for Implementation

93 or available scientific methods are not acceptable, e.g., the scientific methods are deemed
94 unacceptable because they are not clinically validated or are not supported by a valid scientific
95 rationale, FDA may request clinical performance data to support an SE determination. In such
96 cases when clinical data are necessary, it may include, for example, data comparing the
97 technological characteristics of the new device to the predicate device, data assessing whether a
98 change in the indications for use results in a different intended use,¹⁴ or data supporting the
99 assessment of the benefit-risk profile of a new device to demonstrate that the new device is as
100 safe and effective as a predicate device.

101
102 Clinical data provided in support of any marketing submission, including a 510(k) submission,
103 should constitute valid scientific evidence as defined in 21 CFR 860.7(c)(2).¹⁵ Clinical data may
104 include, but are not limited to, results of pre- and post-market clinical investigation(s) of the
105 device (i.e., traditional clinical trials); results of pre- and post-market clinical investigation(s) or
106 other studies reported in the scientific literature of a comparable device; published and/or
107 unpublished reports on clinical experience of either the device in question or a comparable
108 device; and other sources of clinical experience such as registries, adverse event databases, and
109 medical records (e.g., electronic health records, claims).¹⁶ Many of these sources constitute real-
110 world data,¹⁷ and the relevance and reliability of such data should be considered in evaluating
111 whether the data constitutes valid scientific evidence sufficient to support the 510(k) submission.
112 Additionally, when considering whether data collected on a comparable device, such as an
113 earlier version of a device or a similar model of a device, may address certain questions of safety
114 and effectiveness, an adequate justification regarding the applicability of such data should be
115 provided demonstrating why such data would be representative of the new device. In some cases,
116 non-clinical data may also be needed to demonstrate that the devices are comparable and that the
117 clinical data from the comparable device are applicable to the new device. For purposes of this
118 guidance, data obtained from human factors testing is not considered clinical data.

119

is suitable, adequate, validated, and feasible. We will consider if a proposed alternative method could be assessed for equivalency to an animal test method.

¹⁴ As described in the [510\(k\) Program Guidance](#), for purposes of SE, the term “intended use” means the general purpose of the device or its function, and encompasses the indications for use. The term “indications for use” describes the disease or condition that the device will diagnose, treat, prevent, cure, or mitigate, including a description of the patient population for which the device is intended.

¹⁵ 21 CFR 860.7(c)(2) states that “Valid scientific evidence is evidence from well-controlled investigations, partially controlled studies, studies and objective trials without matched controls, well-documented case histories conducted by qualified experts, and reports of significant human experience with a marketed device, from which it can fairly and responsibly be concluded by qualified experts that there is reasonable assurance of the safety and effectiveness of a device under its conditions of use.”

¹⁶ These example sources of clinical data are leveraged from the International Medical Device Regulators Forum Document, “[Clinical Evidence – Key Definitions and Concepts](#),” available at <http://www.imdrf.org/docs/imdrf/final/technical/imdrf-tech-191010-mdce-n55.pdf>.

¹⁷ For additional information regarding real world data, refer to FDA’s guidance, “[Use of Real-World Evidence to Support Regulatory Decision-Making for Medical Devices](#),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/use-real-world-evidence-support-regulatory-decision-making-medical-devices>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

120 **III. Scope**

121 This guidance provides recommendations for when clinical data may be needed to demonstrate
122 that a device reviewed under the 510(k) Program is SE to a predicate device. The
123 recommendations in this guidance are consistent with the [510\(k\) Program Guidance](#) and expand
124 on Section IV.F of that guidance. This guidance also provides additional detail on situations
125 where providing clinical data may be the least burdensome¹⁸ means of demonstrating SE
126 between a new device and a predicate device. FDA developed this guidance to improve the
127 predictability, consistency, and transparency of the 510(k) premarket review process.

128
129 This guidance does not describe situations when postmarket collection of clinical data may be
130 appropriate, such as when clinical data are required in a postmarket surveillance study.¹⁹ This
131 guidance, and the concepts discussed herein, are not intended to propose any changes to
132 applicable statutory and regulatory standards, such as how FDA evaluates SE, or the applicable
133 requirements, including 510(k) content requirements and the requirement for valid scientific
134 evidence.²⁰ This guidance is intended to describe scenarios when clinical data may be necessary
135 and is not intended to supersede applicable regulatory requirements of special controls that
136 outline clinical data requirements for certain device types.

137
138 The principles in this guidance are applicable to devices that are subject to 510(k) review by
139 CDRH and CBER; however, this guidance is not intended to supplant existing device-specific
140 guidance. This guidance does not address review issues unique to combination products. For
141 information on combination products, please refer to the [Office of Combination Products](#)
142 [webpage](#).²¹

143
144 If you have questions about how this guidance and a device-specific guidance apply to a
145 particular issue, we recommend that you consider the general recommendations in this document
146 and discuss specific questions with the appropriate review division associated with your device
147 by submitting a pre-submission. Additional information on the pre-submission program is
148 available in the FDA guidance, “[Requests for Feedback and Meetings for Medical Device](#)
149 [Submissions: The Q-Submission Program](#).”²²

150

¹⁸ See *supra* n. 13.

¹⁹ Section 522 of the FD&C Act. For further information on postmarket surveillance studies, see FDA’s guidance, “[Postmarket Surveillance Under Section 522 of the Federal Food, Drug, and Cosmetic Act](#),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/postmarket-surveillance-under-section-522-federal-food-drug-and-cosmetic-act>.

²⁰ Sections 513(i) and 515 of the FD&C Act, 21 CFR Part 807 Subpart E, and 21 CFR 860.7(c)(2).

²¹ Available at <https://www.fda.gov/combination-products>.

²² Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program>.

151 **IV. Appropriate Use of Clinical Data in 510(k) Decision-**
152 **Making**

153 As described in the [510\(k\) Program Guidance](#), clinical data may be used during the 510(k)
154 review process to support an SE determination at multiple points in the decision tree to address
155 the critical questions in the 510(k) Decision-Making Flowchart.²³

156
157 Typically, clinical data is reviewed after FDA finds that the intended use of the new device and
158 the predicate device are the same, and that the devices have different technological
159 characteristics that do not raise different questions of safety and effectiveness.²⁴ In such cases,
160 clinical data often is used to determine whether the new device is “as safe and effective” as a
161 predicate device. However, clinical data may also be reviewed at other stages of the 510(k)
162 review process. For example, in rare instances, FDA may rely upon clinical data to determine
163 that new²⁵ or modified indications for use fall within the same intended use as a predicate
164 device.²⁶ This guidance describes some of the more common scenarios where clinical data may
165 be necessary to determine SE.

166

167 **V. Scenarios When Clinical Data May be Necessary to**
168 **Determine Substantial Equivalence**

169 FDA initially described the most common scenarios for when clinical data may be necessary in a
170 510(k) submission to demonstrate SE and provided illustrative examples in the [510\(k\) Program](#)
171 [Guidance](#), Section IV.F, “Requests for Performance Data.”

172

173 In this guidance, FDA provides additional clarity on those scenarios (Scenarios 1 – 3 below), and
174 describes another scenario (Scenario 4 below), to provide broad considerations to be used by
175 industry and FDA to help determine whether clinical data may be necessary to demonstrate that a
176 new device is SE to a predicate device:

177

- 178 1. There are differences between the indications for use of the new device and the predicate
179 device, and clinical data may be needed to determine SE.
- 180 2. There are differences between technological characteristics of the new device and the
181 predicate device, and clinical data may be needed to determine SE.
- 182 3. SE between the new device and the predicate device cannot be determined by non-
183 clinical testing (analytical, bench, and/or animal).
- 184 4. A newly identified or increased risk for the predicate device suggests clinical data may be
185 needed for the new device in order to determine SE.

186

²³ See [510\(k\) Program Guidance](#), Appendix A, Decision Points 1 through 4.

²⁴ See *id.* at Decision Points 5a and 5b.

²⁵ As used in this guidance, the term “new” in describing indications for use refers to an indication that is new or differs from that of the predicate device.

²⁶ See *id.* at Decision Point 2.

Contains Nonbinding Recommendations

Draft – Not for Implementation

187 The information below provides additional descriptions of each of the scenarios for when clinical
188 data may be necessary to determine SE, as well as illustrative examples. The applicability of
189 these scenarios may be determined based upon current knowledge, understanding, evidence, and
190 experience available for the new device. Following the least burdensome provisions, the need for
191 clinical data may also change as information on the device type is accrued. FDA acknowledges
192 that there may be situations where one or more of these scenarios exist, but clinical data may not
193 be needed depending on the specific circumstances surrounding the particular new device.
194 Accordingly, for each scenario, FDA has provided examples where clinical data may be needed,
195 as well as examples where clinical data are not typically needed, to determine SE. In addition,
196 there may be other scenarios not described herein for which clinical data may be necessary to
197 determine SE. Note, as described in the [510\(k\) Program Guidance](#), the examples provided below
198 distinguish between examples that are only applicable to diagnostic devices, including in vitro
199 diagnostics (IVDs), and therapeutic devices. This is because there are significant differences in
200 the types of clinical data that may be needed to determine SE for these two categories of devices.
201

202 **A. Scenario #1 – Differences in the indications for use**

203 As described in the [510\(k\) Program Guidance](#), when the indications for use of a new device and
204 predicate device differ, FDA must evaluate whether the indications for use of the new device fall
205 within the same intended use as that of the predicate device. FDA determines the indications for
206 use of the new device based on the proposed labeling²⁷ and the indications for use statement in a
207 510(k). Following review of the proposed labeling and indications for use statement, FDA may
208 rely upon other clinical and/or scientific information submitted with the 510(k) in order to
209 determine if the new device has the same intended use as the predicate device.
210

211 The following factors could impact when clinical data may be necessary to include in a 510(k)
212 submission to demonstrate SE when there are differences between the indications for use of the
213 new device and the predicate device, as shown in illustrative examples 1-A, 1-B, 1-C, and 1-D:
214

- 215 • Differences in the patient population
- 216 • Differences in the disease
- 217 • Differences in the anatomical site, structure, or pathology
- 218 • General to specific considerations²⁸
- 219 • Expansion of the new device’s currently-cleared indications for use
- 220 • Unknown or different benefit-risk profile for the proposed indications for use
- 221

222 Example 1-A: A certain device typically does not require clinical data to be included in a 510(k)
223 submission. However, if a new device is indicated for use in a higher risk population (e.g.,

²⁷ Pursuant to section 513(i)(1)(E)(i) of the FD&C Act, the proposed labeling in a 510(k) submission is used to determine a device’s intended use. The intended use of a device encompasses the indications for use.

²⁸ See 21 CFR 807.92(a)(5); see also FDA’s guidance “[General/Specific Intended Use](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/generalspecific-intended-use-guidance-industry),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/generalspecific-intended-use-guidance-industry>, which identifies the general principles that will be considered by FDA in determining when a specific indication for use is reasonably included within a general indication for use of a medical device for purposes of determining SE.

Contains Nonbinding Recommendations

Draft – Not for Implementation

224 different disease stage) than the predicate device, clinical data may be needed to demonstrate SE
225 if there is increased risk for the use of the new device in the higher risk population due to
226 differences in the benefit-risk profile of the new device for the proposed indications for use when
227 compared to the predicate device.

228

229 Example 1-B: A certain device is indicated for use in a specific anatomic location that is in
230 proximity to critical organs. The manufacturer intends to pursue an indication for use in a
231 different anatomic location that does not represent a new intended use and does not pose
232 additional or different risks. In this scenario, non-clinical data may suffice to demonstrate SE
233 because the indication for use for the predicate device (i.e., the currently-marketed device)
234 represents a higher risk or similar risk scenario than that of the new device. As a result, no
235 additional clinical data are likely to be necessary to demonstrate SE because the benefit-risk
236 profile of the new device with the expanded indications for use is comparable to that of the
237 predicate device.

238

239 Example 1-C: A device is indicated for use in a specific anatomic location, and a manufacturer
240 wants to expand the indications for use to a different anatomic location for the same intended
241 use. There are no other changes to the device. Based on what is known for this device type in the
242 literature and through clinical experience, using the device in this new anatomic location presents
243 an increased risk (for example, due to increased proximity to critical organs or structures, or the
244 indication includes a procedure that is technically more risky or complex). Clinical data may be
245 necessary to demonstrate SE between the new device with the expanded indications for use and
246 the predicate device (i.e., the currently-marketed device) due to the increased risk that may
247 adversely affect the benefit-risk profile of the new device when compared to the predicate
248 device.

249

250 Example 1-D: A predicate laser device is indicated for treatment of a certain skin condition. A
251 new laser device is indicated for treatment of a different skin condition that is not a new intended
252 use. This new laser device utilizes a lower energy wavelength than the predicate device for
253 treatment. Although the lower energy wavelength is not expected to present increased risk
254 compared to the predicate device, clinical data may be needed to demonstrate that the new device
255 has an equivalent benefit-risk profile to the predicate device, given that the new device may
256 result in a different degree of benefit for treatment compared to the predicate device due to both
257 the lower energy wavelength and the difference in skin conditions.

258

259 **B. Scenario #2 – Differences in the technological** 260 **characteristics**

261 As discussed in the [510\(k\) Program Guidance](#), clinical data may be necessary to include in a
262 510(k) submission when there are differences between the technological characteristics of the
263 new device and the predicate device that do not raise different questions of safety and
264 effectiveness in order to establish that a new device performs equivalent to the predicate device
265 despite the differences in those characteristics.²⁹

²⁹ Section 513(i)(1)(A)(ii) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

266
267 The following factors should be considered in determining whether clinical data may be
268 necessary to include in a 510(k) submission to demonstrate SE when there are differences
269 between the technological characteristics^{30,31} of the new device and the predicate device, as
270 illustrated in examples 2-A, 2-B, and 2-C:

- 271
- 272 • Significant change in materials
 - 273 • Significant change in device design
 - 274 • Significant change in energy source
 - 275 • Significant change in other device features
- 276

277 Example 2-A: An implanted device is used to provide anatomic support resulting in improved
278 function. Most such devices are made of non-resorbable materials. Available performance data
279 on such devices may not be applicable to a device comprised of resorbable material, which
280 resorbs in vivo over time. For this difference in technology, assuming it does not raise different
281 questions of safety and effectiveness, clinical data may be needed to support the SE
282 determination.

283

284 Example 2-B: An IVD uses a monoclonal antibody as a critical reagent. If the manufacturer
285 decides to change to a different clone from the previous antibody, clinical data may be needed to
286 support the SE determination, as the differences in technological characteristics between the new
287 IVD and the predicate IVD raise a question of whether the clinical performance of the new IVD
288 can be expected to be equivalent to the clinical performance of the predicate IVD.

289

290 Example 2-C: A manufacturer chooses to add additional sizes of an implanted device to its
291 existing line of cleared, implanted devices. No other changes are made to the design, materials,
292 or other device features. The new sizes are within the minimum and maximum of the cleared,
293 implanted devices of the device type. The new devices can likely be assessed using adequate
294 non-clinical testing methods to determine SE to the predicate device. It is unlikely that clinical
295 data would be necessary to evaluate this change in technological characteristic.

296

297 However, if the size of the new implanted device would become the new maximum or minimum
298 size of all cleared, implanted devices of the device type, expanding the range of device sizes,
299 provided that the intended use of the new device is the same as the predicate device, clinical data
300 may be needed to support an SE determination.

301

³⁰ “Differences in technological characteristics” is defined in section 513(i)(1)(B) of the FD&C Act.

³¹ The [510\(k\) Program Guidance](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k) describes overarching aspects for consideration regarding device design, materials, energy source, and other key technological features. For further information, please see Section IV.E of the [510\(k\) Program Guidance](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k), available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k>.

302 **C. Scenario #3 – SE cannot be determined by non-clinical**
303 **testing**

304 Clinical data may be necessary to include in a 510(k) submission when non-clinical testing, such
305 as analytical, bench, and/or animal testing, is not adequate to establish that the new device is SE
306 to the predicate device.³²

307
308 The following factors represent considerations for determining when clinical data may be
309 necessary to demonstrate SE as non-clinical testing may not be appropriate for a particular
310 device, because:

- 311
- 312 • There is no model (e.g., analytical, bench, animal) available
 - 313 • The available model(s) may not be adequate because the model has certain limitations
314 that do not allow for an adequate assessment
 - 315 • The model may not be predictive of clinical outcomes
 - 316 • There are anatomical and/or pathophysiological species-specific questions that rely on
317 clinical evidence
- 318

319 Example 3-A: For a new device with an intended use for treatment of schizophrenia, clinical data
320 may be needed to demonstrate that the device is SE to the predicate device given the limited
321 availability of non-clinical models and inadequate predictions of clinical outcomes for
322 schizophrenia.

323
324 Example 3-B: A basic medical image management and processing system that adds a new organ-
325 specific processing, filtering, or enhancement feature may need to submit clinical data to
326 demonstrate that the new device is SE to the predicate device. This may occur in scenarios where
327 there is no phantom that accurately models the organ in that imaging modality, so we
328 recommend using clinical images as part of the device evaluation.

329
330 Example 3-C: For a device intended for use to support hemostasis, clinical data may be
331 necessary to demonstrate that the new device is SE to the predicate device given the inadequacy
332 of current bench and animal models to be predictive and representative of human performance
333 due to the differences in coagulation pathways between animals and humans.

334
335 Example 3-D: For a device intended for use in screening donors of blood and blood products for
336 transfusion-transmitted infections, clinical data may be necessary to demonstrate that the new
337 device is SE to the predicate device given the inability of analytical testing to evaluate the
338 clinical performance of the assay and the risks to the blood supply associated with incorrect
339 results.

340
341 Example 3-E: For IVDs, including for an IVD intended for point-of-care use where the predicate
342 device is not intended for point-of-care use, clinical data may be necessary to demonstrate that
343 the new device performs equivalent and is SE to the predicate device. This is due to multiple

³² Section 513(i)(1)(A) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

344 factors, including the variety of clinical environments and the diverse populations with which the
345 device is intended to be used, which can affect the performance of the device and cannot be
346 evaluated solely through analytical data.

347
348 Example 3-F: For a device intended for an aesthetic purpose (e.g., to treat acne scars, or to
349 reduce wrinkles), it may be difficult to demonstrate the effectiveness of the new device solely
350 with non-clinical (e.g., animal) data to determine SE to a predicate device. This is because there
351 are no appropriate animal models for device types intended for aesthetic purposes, and validated
352 aesthetic measures of effectiveness and the translatability of such measures in humans have not
353 been established. For this reason, in many scenarios, clinical data may be necessary to
354 demonstrate that a new device intended for an aesthetic purpose is SE to the predicate device.

355

356 **D. Scenario #4 – Newly identified or increased risk for the**
357 **predicate device**

358 Although significant attention is applied to the design, testing, manufacturing, and evaluation of
359 medical devices prior to their introduction into the marketplace, not all information regarding
360 benefits and risks is available nor can be generally known at that time. New information about a
361 device’s safety, including unexpected adverse events, may become available once the device is
362 more widely distributed and used in clinical practice.

363

364 In such cases, there may not be identified differences between the technological characteristics of
365 new device and the predicate device that raise different questions of safety or effectiveness.
366 However, there may be an awareness of new scientific information regarding a newly identified
367 or increased risk of the predicate device, and clinical data may be needed to determine SE in
368 light of the new scientific information.

369

370 As described in the [510\(k\) Program Guidance](#) (Section IV.F), new scientific information may
371 affect FDA’s expectations concerning the type and level of performance data to be included in a
372 510(k) submission. FDA may learn of these new or increased risks for a device (compared to
373 what was known prior to introduction into the marketplace) from voluntarily-reported adverse
374 events or literature, or from other sources of real-world data (e.g., 522 postmarket surveillance
375 studies^{33,34} or recalls³⁵), and incorporate that information into its review of premarket
376 submissions and SE determinations. Information regarding new or increased risks for a device is
377 often publicly communicated (e.g., via safety communication, guidance, advisory committee
378 meeting) by FDA. When requesting clinical data during premarket review due to a new or

³³ For further information on postmarket surveillance studies, see FDA’s guidance, “[Postmarket Surveillance Under Section 522 of the Federal Food, Drug, and Cosmetic Act](#),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/postmarket-surveillance-under-section-522-federal-food-drug-and-cosmetic-act>.

³⁴ See FDA’s webpage on the 522 Postmarket Surveillance Studies Program, available at <https://www.fda.gov/medical-devices/postmarket-requirements-devices/522-postmarket-surveillance-studies>.

³⁵ See FDA’s webpage on Recalls, Market Withdrawals, & Safety Alerts available at <https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts>, as well as the Medical Device Recalls database, available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRES/res.cfm>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

379 increased risk, FDA intends to provide an explanation of the reason(s) for the request and why
380 such information is necessary to determine whether the new device is SE, consistent with the
381 FDA guidance on “[Developing and Responding to Deficiencies in Accordance with the Least](#)
382 [Burdensome Provisions](#).”³⁶

383
384 Whenever possible, FDA recommends that manufacturers should not use certain devices as
385 predicate devices if they exhibit new or increased risks, especially if an alternative predicate
386 device exists without the new or increased risk.³⁷ However, in certain circumstances, there may
387 not be an alternative predicate device available without the new or increased risk. Devices that
388 exhibit new or increased risks may lead FDA to consider the need for additional data, such as
389 clinical data, in the premarket submissions for such technology, as illustrated in examples 4-A,
390 4-B, and 4-C.

391
392 Example 4-A: Through review of recalls, voluntarily-reported adverse events, and published
393 scientific literature, FDA became aware of certain malfunctions for a particular device. However,
394 based on FDA’s assessment of the totality of clinical (e.g., published medical literature) and non-
395 clinical data (e.g., non-clinical bench performance testing), FDA determined that detailed non-
396 clinical testing, accompanied by appropriate instructions for use, could adequately demonstrate
397 whether the risk for the new device was adequately mitigated by its design and technological
398 features. FDA determined that additional clinical data was not necessary to demonstrate SE for
399 new 510(k) submissions that may use this device as the predicate, provided that the appropriate
400 non-clinical testing and certain labeling considerations are addressed in the 510(k) submission.

401
402 Example 4-B: A device was initially cleared by FDA without the inclusion of clinical data in the
403 510(k). Following introduction of the device into the marketplace, recalls and other postmarket
404 surveillance data reviewed by FDA suggested safety concerns related to component failure in the
405 device. After a thorough review of the available data, FDA issued a class-wide postmarket
406 surveillance study order under section 522 of the FD&C Act for currently marketed devices for
407 this device type and began requesting that clinical data be included in 510(k) submissions for
408 new devices seeking marketing clearance for this device type to ensure an adequate safety profile
409 prior to marketing.

410
411 Example 4-C: There was a device issue reported that could lead to significant patient injury in
412 surgical procedures. For this cleared device, the primary evidence demonstrating SE in 510(k)
413 submissions had been non-clinical design verification and validation testing of the technological
414 characteristics of the device. The manufacturer voluntarily recalled the device, submitted a new
415 510(k) to address the issue, and included non-clinical and clinical performance data because the
416 changes could significantly affect the safety or effectiveness of the device.³⁸ FDA issued device-

³⁶ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/developing-and-responding-deficiencies-accordance-least-burdensome-provisions>.

³⁷ Please see FDA’s draft guidance, “[Best Practices for Selecting a Predicate Device to Support a Premarket Notification \[510\(k\)\] Submission](#),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/best-practices-selecting-predicate-device-support-premarket-notification-510k-submission>. When final, that guidance will represent FDA’s current thinking on that topic.

³⁸ See 21 CFR 807.81.

Contains Nonbinding Recommendations

Draft – Not for Implementation

417 specific guidance to outline recommendations for non-clinical and clinical performance testing
418 for this device type.

DRAFT