



May 31, 2024

Aiiso Hx Yang
COO & VP
CorDx, Inc.
9540 Waples Street, #C
San Diego, CA 92121

Re: EUA240006/S001
Trade/Device Name: CorDx TyFast Flu A/B & COVID-19 At Home Multiplex Rapid Test
Dated: May 20, 2024
Received: May 20, 2024

Dear Aiiso Hx Yang:

This is to notify you that your request to update the CorDx TyFast Flu A/B & COVID-19 At Home Multiplex Rapid Test to extend the shelf-life expiration date to 12 months when stored at 2°C – 30°C, based on the results of your ongoing stability studies, is granted. Upon review, we concur that the data and information submitted in EUA240006/S001 supports the requested update for the CorDx TyFast Flu A/B & COVID-19 At Home Multiplex Rapid Test. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the CorDx TyFast Flu A/B & COVID-19 At Home Multiplex Rapid Test issued on April 5, 2024.

Sincerely yours,

Uwe Scherf, M.Sc., Ph.D.
Director, Division of Microbiology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health