



July 1, 2024

Daniel Hsu
Director of Compliance and Technology
Advin Biotech Inc.
10237 Flanders Ct.
San Diego, CA, 92121

Re: EUA220358/S013
Trade/Device Name: Advin COVID-19 Antigen Test @Home
Dated: June 3, 2024
Received: June 3, 2024

Dear Daniel Hsu:

This is to notify you that your request to update the Advin COVID-19 Antigen Test @Home to extend the shelf-life expiration date to 24 months when stored at 2°C – 30°C, based on the results of your ongoing stability studies, is granted. Upon review, we concur that the data and information submitted in EUA220358/S013 supports the requested update for the Advin COVID-19 Antigen Test @Home. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the Advin COVID-19 Antigen Test @Home issued on December 22, 2022.

Sincerely yours,

Uwe Scherf, M.Sc., Ph.D.
Director, Division of Microbiology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health